

This memo provides responsive information to inquiries on the safety of SCUBA diving in patients with implanted Titan® Inflatable Penile Protheses (IPP), Titan Touch IPP, or Torosa® Testicular Protheses. It includes relevant data from approved product labeling, testing data, and published literature.

Product Labeling

Please refer to the Titan IPP Instructions for Use (IFU) and Torosa Prothesis IFU for information about approved uses and important safety information.^{1,2} There are no warnings, precautions, or contraindications in Titan labeling or Torosa labeling about high pressure environments. The Titan IFU contains the following warning about movement and activity: "Patients must be warned that any movements or activities that put pressure directly on the pump or deflate mechanism can cause involuntary erection or deflation. This includes, but is not limited to straddle seating, such as on motorcycles, bicycles, horses, etc. Obesity may also be a contributing factor for stresses and pressure."

Published Literature

Search Strategy

The Coloplast Medical Affairs team performed a literature search in PubMed in February 2025. The search strategy included terms for Titan, penile prosthesis, Torosa, testicular implant, and diving. It excluded nonhuman studies, non-English publications, and publications older than 10 years. Additional relevant publications were also evaluated for content and references.

Search Results

The search did not identify any results.

If you need additional information, please contact IUMedicalAffairs@Coloplast.com.

Disclaimers

Medical Information

This information is intended to inform clinical decision making, support the safe and effective use of Coloplast products, and foster scientific exchange. It is not intended to be medical advice or offer recommendations inconsistent with product labeling.

Limitations of Medical Research

The information derived from our literature searches should not be considered comprehensive, as medical literature databases are inherently selective.

How to Report an Adverse Event

To report a patient adverse event or product complaint, please submit information to your local Coloplast representative or contact complaints@coloplast.com.

References

1. Titan and Titan Touch Instructions for Use (North America) 70003006 Rev. A. Humlebaek, DK: Coloplast A/S.; July 2024.

2. Torosa Saline-Filled Testicular Protheses Instructions for Use (North America)
70002200 Rev. B. Humlebaek, DK: Coloplast A/S; October 2021.
3. Coloplast. Titan data on file. VV-0342048 v1.0.